INTOXILYZER 5000 and 5000 KB INSPECTION

Serial No. 68-010634 Location TOX LAB

		Check when done
1.	Is three pronged grounded outlet used?	✓
2.	Is the warm-up time less than 20 minutes?	✓
3.	Is breath tube heated?	✓
4.	Diagnostic check complete?	✓
5.	DVM (Switch 2 Up) Value —	NA
6.	Time, date, year correct? ReSet if necessary.	✓
7.	Print test. (Mode P). (Attach test record)	✓

8. Calibration Check. Use Simulator Standard Solution. Do three sets in ACA mode.(Attach test records)

Sim.Serial No. 69986 Sim. Soln. No. 477 Sim. Temp. 34.0°c

Test 1	0.115	Test 4	0.115	Test 7	0.114
Test 2	0.116	Test 5	0.115	Test 8	0.114
Test 3	0.115	Test 6	0.115	Test 9	0.115

Average 0.115 AC

9.	Low AC. Use 0.02 AC in ACA mode (Attach test record)	✓
10.	Linearity Test. Use 0.30 AC in ACA mode. (Attach test record)	✓
11.	Interferent check. Use 0.05% Actone plus 0.10 AC ethanol in ABA mode. (Attach test record)	✓
12.	RFI check. Run CMS mode. Key radio on 1st room air. (Attach test record)	✓
13.	CMS mode. Use standard simulator solution for B, C, & B of CMS. (Attach test record)	✓

14. Remarks/Maintenance record:

2 April 2002
Date

Margaret A Pearson
Field Inspector's Signature

Form 110-I(3/02)

F:\Margy\InspectIntox5000&5000KB.wpd

MAP

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	11:32 CST
02 STD. SOL.	.299	11:32 CST
03 ROOM AIR	.000	11:33 CST
04 STD. SOL.	.299	11:33 CST
05 ROOM AIR	.000	11:33 CST
06 STD. SOL.	.300	11:34 CST
07 ROOM AIR	.000	11:34 CST

08 SIM TEMP=33.9 LOCATION=TOXL

SIMUL SER NO=G9917
STD SOL NO=292
COUNTY=08 OPER NO =107501

☐ 20 minute waiting period was ascertained

☐ A clean mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☐ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

Linearity 0.292 AC

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	11:25 CST
02 STD. SOL.	.020	11:26 CST
03 ROOM AIR	.000	11:26 CST
04 STD. SOL.	.020	11:26 CST
05 ROOM AIR	.000	11:27 CST
06 STD. SOL.	.019	11:27 CST
07 ROOM AIR	.000	11:27 CST

08 SIM TEMP=33.9 LOCATION=TOXL

SIMUL SER NO=G9234
STD SOL NO=021
COUNTY=08 OPER NO =107501

☐ 20 minute waiting period was ascertained

☐ A clean mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☐ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

LOW AC 0.021 AC.

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	45±5	DEG C
02 ROOM AIR	.000	11:02 CST
03 SUBJECT TEST 1	.107	11:02 CST
04 ROOM AIR	.000	11:04 CST
05 STD. SOL.	.111	11:06 CST
06 ROOM AIR	.000	11:07 CST
07 SUBJECT TEST 2	.107	11:08 CST
08 ROOM AIR	.000	11:08 CST
09 REPORTED AC	.10	11:02 CST

DIFFERENCE OK

NO RFI DETECTED
10 SIM TEMP=34.0 LOCATION=TOXL

SUB NAME=CMS,NA,NA
SUB DOB =00/00/00
SUB SEX =FEMALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501
ALL BOXES CHECKED AT COMPLETION OF
STEP AND TEST REC SIGNED Y/N? Y

☑ 20 minute waiting period was ascertained

☒ A clean mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☒ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

CMS

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	10:57 CST
02 SUBJECT TEST 1	.093	10:58 CST
03 INTERFERENT SUBTRACTED		
04 ROOM AIR	.000	10:58 CST
05 REPORTED AC	.09	10:58 CST

NO RFI DETECTED
06 SIM TEMP=??.? LOCATION=TOXL

SUB NAME=INTERFERENT,NA,NA
SUB DOB =00/00/00
SUB SEX =FEMALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501
ALL BOXES CHECKED AT COMPLETION OF
STEP AND TEST REC SIGNED Y/N?

☐ 20 minute waiting period was ascertained

☒ A clean mouthpiece was used for and disposed of after the first subject test

☐ Standard Solution test completed

☐ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

INTERFERENT

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	45±5	DEG C
02 ROOM AIR	.000	10:55 CST

SUB NAME=RFI,NA,NA
SUB DOB =00/00/00
SUB SEX =FEMALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501
INVALID TEST
INHIBITED - RFI

☐ ~~20 minute~~ waiting period was ascertained

☐ ~~A clean~~ mouthpiece was used for and disposed of after the first subject test

☐ ~~Standard~~ Solution test completed

☐ ~~A clean~~ mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

RFI

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	10:48 CST
02 STD. SOL.	.114	10:48 CST
03 ROOM AIR	.000	10:48 CST
04 STD. SOL.	.114	10:49 CST
05 ROOM AIR	.000	10:49 CST
06 STD. SOL.	.115	10:49 CST
07 ROOM AIR	.000	10:49 CST

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501

☐ ~~20 minute~~ waiting period was ascertained

☐ ~~A clean~~ mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☐ ~~A clean~~ mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

ACA

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	10:42 CST
02 STD. SOL.	.115	10:42 CST
03 ROOM AIR	.000	10:42 CST
04 STD. SOL.	.115	10:43 CST
05 ROOM AIR	.000	10:43 CST
06 STD. SOL.	.115	10:43 CST
07 ROOM AIR	.000	10:44 CST

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501

☐ ~~20 minute~~ waiting period was ascertained

☐ ~~A clean~~ mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☐ ~~A clean~~ mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

ACA

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/02/2002 07/00

TEST	AC	TIME
01 ROOM AIR	.000	10:38 CST
02 STD. SOL.	.115	10:38 CST
03 ROOM AIR	.000	10:38 CST
04 STD. SOL.	.116	10:38 CST
05 ROOM AIR	.000	10:39 CST
06 STD. SOL.	.115	10:39 CST
07 ROOM AIR	.000	10:39 CST

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=G9986
STD SOL NO=477
COUNTY=08 OPER NO =107501

☐ ~~20 minute waiting~~ period was ascertained

☐ ~~A clean mouth~~piece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☐ ~~A clean mouth~~piece was used for and disposed of after the second subject test

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

ACA

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

SN 69-010684 VERSION GI286.08
SLAVE 75_2240
DATE 04/02/2002 TIME 10:36 CST

ABCDEFGHIJKLMNOPQRSTUVWXYZ0123
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
ABCDEFGHIJKLMNO
ABCDEFGHIJKLMNOPQR
ABCDEFGHIJKLMNOPQRSTUV
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
** CURRENT INSTRUMENT SETUP **
SETUP WORD = 0x00B9
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = TRUE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
REPRINT CRD? Y/N = TRUE

INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE SETUP

☐ 20 minute waiting period was ascertained

☐ A clean mouthpiece was used for and disposed of after the first subject test

☐ Standard Solution test completed

☐ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Margy Pearson

Remarks:

Print Test

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY