INTOXILYZER 5000 KB AND 5000 KB-EP INSPECTION

✓ Database emm 7-15-08

Serial No.: 68-010634 Location: TOKL

		Check When Done
1.	Is a three-pronged grounded outlet used?	✓
2.	Is the warm-up time less than 20 minutes?	✓
3.	Is the breath tube heated?	✓
4.	Is the diagnostic check complete?	✓
5.	Is the time, date, and year correct (reset if necessary)?	✓
6.	Print test. (Attach test record.)	✓

7. Calibration Check. Use Simulator Standard Solution. Do three sets in ACA mode. (Attach test record.)

Sim. Serial No. DR 3804 Sim. Soln. No. 549 Sim. Temp. 34.0°C

Test 1 0.106 Test 4 0.107 Test 7 0.107

Test 2 0.107 Test 5 0.107 Test 8 0.107

Test 3 0.107 Test 6 0.107 Test 9 0.107

Average 0.107

8.	Low AC. Use 0.015 AC in ACA mode. (Attach test record.)	✓
9.	Linearity Test. Use 0.30 AC in ACA mode. (Attach test record.)	✓
10.	Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.)	✓
11.	RFI check. Run CMS mode. Key radio on first room air. (Attach test record.)	✓
12.	CMS mode. Use standard simulator solution for B, C, & B of CMS. (Attach test record.)	

13. Remarks/Maintenance record: Recalibrated, Eprom change (1286-32) Annual inspection complete.

Date: 22 Apr 2008 Field Inspector's Signature: Deb Kashur

Form 110-I (4/03)

SN 68-010634 VERSION G1286.32
SLAVE 75_2240
DATE 04/22/2008 TIME 15:46 CDT

ABCDEFGHIJKLMNOPQRSTUVWXYZ0123
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
ABCDEFGHIJKLMNO
ABCDEFGHIJKLMNOPQR
ABCDEFGHIJKLMNOPQRSTUV
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
** CURRENT INSTRUMENT SETUP **
SETUP WORD = 0x00B9
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = TRUE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
REPRINT CRD? Y/N = TRUE

INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE SETUP

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☐ ~~Standard Solution test completed~~

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:

Print test

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	15:47 CDT
02 STD. SOL.	.106	15:47 CDT
03 ROOM AIR	.000	15:47 CDT
04 STD. SOL.	.107	15:47 CDT
05 ROOM AIR	.000	15:48 CDT
06 STD. SOL.	.107	15:48 CDT
07 ROOM AIR	.000	15:48 CDT

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=DR3804
STD SOL NO=549
COUNTY=08 OPER NO =777777

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:

ACA Calibration check

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	15:50 CDT
02 STD. SOL.	.107	15:50 CDT
03 ROOM AIR	.000	15:50 CDT
04 STD. SOL.	.107	15:51 CDT
05 ROOM AIR	.000	15:51 CDT
06 STD. SOL.	.107	15:51 CDT
07 ROOM AIR	.000	15:51 CDT

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=DR3804
STD SOL NO=549
COUNTY=08 OPER NO =777777

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:
ACA Calibration check

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	15:53 CDT
02 STD. SOL.	.107	15:53 CDT
03 ROOM AIR	.000	15:53 CDT
04 STD. SOL.	.107	15:53 CDT
05 ROOM AIR	.000	15:54 CDT
06 STD. SOL.	.107	15:54 CDT
07 ROOM AIR	.000	15:54 CDT

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=DR3804
STD SOL NO=549
COUNTY=08 OPER NO =777777

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☐ ~~Standard Solution test completed~~

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:
ACA Calibration check

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	15:56 CDT
02 STD. SOL.	.014	15:57 CDT
03 ROOM AIR	.000	15:57 CDT
04 STD. SOL.	.013	15:57 CDT
05 ROOM AIR	.000	15:57 CDT
06 STD. SOL.	.013	15:58 CDT
07 ROOM AIR	.000	15:58 CDT

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=DR3454
STD SOL NO=016
COUNTY=08 OPER NO =777777

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:

Low AC

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist - Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	16:06 CDT
02 STD. SOL.	.292	16:06 CDT
03 ROOM AIR	.000	16:06 CDT
04 STD. SOL.	.290	16:06 CDT
05 ROOM AIR	.000	16:07 CDT
06 STD. SOL.	.291	16:07 CDT
07 ROOM AIR	.000	16:07 CDT

08 SIM TEMP=34.0 LOCATION=TOXL

SIMUL SER NO=G8115
STD SOL NO=298
COUNTY=08 OPER NO =777777

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:

Linearity test

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist - Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 ROOM AIR	.000	16:10 CDT
02 SUBJECT TEST 1	.094	16:10 CDT
03 INTERFERENT SUBTRACTED		
04 ROOM AIR	.000	16:11 CDT
05 REPORTED AC	.09	16:10 CDT

DIFFERENCE OK

NO RFI DETECTED
06 SIM TEMP=??.? LOCATION=TOXL

SUB NAME=NA,NA,NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=DR5191
STD SOL NO=100
COUNTY=08 OPER NO =777777
ALL BOXES CHECKED AT COMPLETION OF
STEP AND TEST REC SIGNED Y/N? ?

☒ 20 minute waiting period was ascertained

☒ A clean mouthpiece was used for and disposed of after the first subject test

☐ ~~Standard Solution test completed~~

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:
Interferent check

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist - Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	4515	DEG C
02 ROOM AIR	.000	16:19 CDT

SUB NAME=NA,NA,NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=00000
STD SOL NO=000
COUNTY=08 OPER NO =777777
INVALID TEST
INHIBITED - RFI

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☐ ~~Standard Solution test completed~~

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:
RFI check

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist - Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
04/22/2008 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	45±5	DEG C
02 ROOM AIR	.000	16:22 CDT
03 SUBJECT TEST 1	.104	16:22 CDT
04 ROOM AIR	.000	16:23 CDT
05 STD. SOL.	.106	16:24 CDT
06 ROOM AIR	.000	16:25 CDT
07 SUBJECT TEST 2	.101	16:27 CDT
08 ROOM AIR	.000	16:27 CDT
09 REPORTED AC	.10	16:27 CDT

DIFFERENCE OK

NO RFI DETECTED
10 SIM TEMP=34.0 LOCATION=TOXL

SUB NAME=NA, NA, NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=DR3804
STD SOL NO=549
COUNTY=08 OPER NO =777777
ALL BOXES CHECKED AT COMPLETION OF
STEP AND TEST REC SIGNED Y/N? Y

☒ 20 minute waiting period was ascertained

☒ A clean mouthpiece was used for and disposed of after the first subject test

☒ Standard Solution test completed

☒ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature Deb Kashur

Remarks:

CMS

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY