INTOXILYZER 5000 KB AND 5000 KB-EP INSPECTION

Serial No.: 68-010634 Location: Sargent Co SO

		Check When Done
1.	Is a three-pronged grounded outlet used?	✓
2.	Is the warm-up time less than 20 minutes?	✓
3.	Is the breath tube heated?	✓
4.	Is the diagnostic check complete?	✓
5.	Is the time, date, and year correct (reset if necessary)?	✓
6.	Print test. (Attach test record.)	✓

7. Calibration Check. Use Simulator Standard Solution. Do three sets in ACA mode. (Attach test record.)

Sim. Serial No. D23454 Sim. Soln. No. 500 Sim. Temp. 34.0

Test 1 .108 Test 4 .109 Test 7 .110

Test 2 .108 Test 5 .109 Test 8 .110

Test 3 .109 Test 6 .110 Test 9 .110

Average .109

8.	Low AC. Use 0.015 AC in ACA mode. (Attach test record.)	✓
9.	Linearity Test. Use 0.30 AC in ACA mode. (Attach test record.)	✓
10.	Interferent check. Use 0.05% Acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.)	✓
11.	RFI check. Run CMS mode. Key radio on first room air. (Attach test record.)	✓
12.	CMS mode. Use standard simulator solution for B, C, & B of CMS. (Attach test record.)	✓

13. Remarks/Maintenance record: Annual Inspection

Date: 5-19-04 Field Inspector's Signature: [signature]

Rev Map

Form 110-I (4/03)

SN 68-010634 VERSION G1286.08
SLAVE 75_2240
DATE 05/19/2004 TIME 12:27 CDT

ABCDEFGHIJKLMNOPQRSTUVWXYZ0123
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
ABCDEFGHIJKLMNO
ABCDEFGHIJKLMNOPQR
ABCDEFGHIJKLMNOPQRSTUV
ABCDEFGHIJKLMNOPQRSTUVWXYZ0123456789!@#$abcde
** CURRENT INSTRUMENT SETUP **
SETUP WORD = 0x00B9
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = TRUE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
REPRINT CRD? Y/N = TRUE

INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE SETUP

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☐ ~~Standard Solution test completed~~

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature [signature]

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:37 CDT
02 STD. SOL.	.108	12:37 CDT
03 ROOM AIR	.000	12:38 CDT
04 STD. SOL.	.108	12:38 CDT
05 ROOM AIR	.000	12:38 CDT
06 STD. SOL.	.109	12:39 CDT
07 ROOM AIR	.000	12:39 CDT

08 SIM TEMP=34.0 LOCATION=FORM

SIMUL SER NO=DR3454
STD SOL NO=500
COUNTY=41 OPER NO =161616

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature [signature]

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:41 CDT
02 STD. SOL.	.109	12:41 CDT
03 ROOM AIR	.000	12:41 CDT
04 STD. SOL.	.109	12:42 CDT
05 ROOM AIR	.000	12:42 CDT
06 STD. SOL.	.110	12:42 CDT
07 ROOM AIR	.000	12:42 CDT

08 SIM TEMP=34.0 LOCATION=FORM

SIMUL SER NO=DR3454
STD SOL NO=500
COUNTY=41 OPER NO =161616

- [x] ~~20 minute waiting period was ascertained~~
- [ ] ~~A clean mouthpiece was used for and disposed of after the first subject test~~
- [x] Standard Solution test completed
- [ ] ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature [signature]

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:45 CDT
02 STD. SOL.	.110	12:45 CDT
03 ROOM AIR	.000	12:46 CDT
04 STD. SOL.	.110	12:46 CDT
05 ROOM AIR	.000	12:46 CDT
06 STD. SOL.	.110	12:46 CDT
07 ROOM AIR	.000	12:47 CDT

08 SIM TEMP=34.0 LOCATION=FORM

SIMUL SER NO=DR3454
STD SOL NO=500
COUNTY=41 OPER NO =161616

- [x] ~~20 minute waiting period was ascertained~~
- [ ] ~~A clean mouthpiece was used for and disposed of after the first subject test~~
- [x] Standard Solution test completed
- [ ] ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature [signature]

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:29 CDT
02 STD. SOL.	.013	12:30 CDT
03 ROOM AIR	.000	12:30 CDT
04 STD. SOL.	.013	12:30 CDT
05 ROOM AIR	.000	12:30 CDT
06 STD. SOL.	.013	12:31 CDT
07 ROOM AIR	.000	12:31 CDT

08 SIM TEMP=34.0 LOCATION=FORM

SIMUL SER NO=DR3453
STD SOL NO=015
COUNTY=41 OPER NO =161616

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:33 CDT
02 STD. SOL.	.283	12:33 CDT
03 ROOM AIR	.000	12:34 CDT
04 STD. SOL.	.287	12:34 CDT
05 ROOM AIR	.000	12:34 CDT
06 STD. SOL.	.287	12:35 CDT
07 ROOM AIR	.000	12:35 CDT

08 SIM TEMP=34.1 LOCATION=FORM

SIMUL SER NO=DR3374
STD SOL NO=293
COUNTY=41 OPER NO =161616

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the Instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 ROOM AIR	.000	12:49 CDT
02 SUBJECT TEST	.092	12:49 CDT
03 INTERFERENT SUBTRACTED		
04 ROOM AIR	.000	12:50 CDT
05 REPORTED AC	.09	12:49 CDT

NO RFI DETECTED
06 SIM TEMP=??.? LOCATION=FORM

SUB NAME=INTERFERENT, CHECK, NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=DR3451
STD SOL NO=098
COUNTY=41 OPER NO =161616

☑ 20 minute waiting period was ascertained

☑ A clean mouthpiece was used for and disposed of after the first subject test

☑ Standard Solution test completed

☑ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	45±5	DEG C
02 ROOM AIR	.000	12:53 CDT

SUB NAME=RFI, CHECK, NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=DR3454
STD SOL NO=500
COUNTY=41 OPER NO =161616
INVALID TEST
INHIBITED - RFI

☑ 20 minute waiting period was ascertained

☑ A clean mouthpiece was used for and disposed of after the first subject test

☑ Standard Solution test completed

☑ A clean mouthpiece was used for and disposed of after the second subject test

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY

CMI INC
INTOXILYZER - ALCOHOL ANALYZER
ND MODEL 5000 SN 68-010634
05/19/2004 07/00

TEST	AC	TIME
01 DIAGNOSTIC OK	45±5	DEG C
02 ROOM AIR	.000	12:55 CDT
03 SUBJECT TEST 1	.106	12:56 CDT
04 ROOM AIR	.000	12:57 CDT
05 STD. SOL.	.109	12:58 CDT
06 ROOM AIR	.000	13:00 CDT
07 SUBJECT TEST 2	.105	13:01 CDT
08 ROOM AIR	.000	13:01 CDT
09 REPORTED AC	.10	13:01 CDT

DIFFERENCE OK

NO RFI DETECTED
10 SIM TEMP=34.0 LOCATION=FORM

SUB NAME=CMS, TEST, NA
SUB DOB =00/00/00
SUB SEX =MALE WEIGHT =000
TEST=OTH CIT=NA
DR. LIC.=NA/NA
SIMUL SER NO=DR3454
STD SOL NO=500
COUNTY=41 OPER NO =161616
ALL BOXES CHECKED AT COMPLETION OF
STEP AND TEST REC SIGNED Y/N? Y

☐ ~~20 minute waiting period was ascertained~~

☐ ~~A clean mouthpiece was used for and disposed of after the first subject test~~

☒ Standard Solution test completed

☐ ~~A clean mouthpiece was used for and disposed of after the second subject test~~

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator's Signature

Remarks:

Form 106-KB
(8-99)

INTOXILYZER TEST RECORD AND CHECKLIST
State Toxicologist • Bismarck, ND 58501

OPERATOR COPY