State of North Dakota)
)ss
County of Burleigh)

I, Margaret A. Pearson, do hereby certify that I am a duly-appointed State Toxicologist for the State of North Dakota and an official custodian of the records and files of the office thereof, that I have carefully compared the

Ethanol Breath Standard Analytical Report, Lot No. 659358 (September 3, 2009)

hereto attached with the respective original as the same appears of record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, and find the same to be a true and correct copy thereof and of the whole thereof. In witness whereof I have set my hand at the city of Bismarck, in said county this:

3 day of September, 2009

Margaret A Pearson
Margaret A. Pearson, State Toxicologist

State of North Dakota)
)ss
County of Burleigh)

On this 3 day of September, 2009, before me personally appeared Margaret A. Pearson, known to me to be a State Toxicologist for the State of North Dakota, and acknowledged to me that he/she has executed the same.

Subscribed to and sworn before me this:

3 day of September, 2009

CINDY LEINGANG
Notary Public
State of North Dakota
My Commission Expires Jan. 11, 2011

Cindy Leingang
Cindy Leingang, Notary Public, State of North Dakota
My Commission Expires January 11, 2011

(SEAL)

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
Form: BrF-112- G

ETHANOL BREATH STANDARD ANALYTICAL REPORT

Ethanol Breath Standard Lot Number **659358** Expiration Date **September 3, 2011**

This standard was analyzed by CALGAZ with reported results to be: **208.4** ppm which is the equivalent of 0.080 AC of Ethanol in Nitrogen. CALGAZ has provided a Certificate of Analysis traceable to N.I.S.T. RGM Ethanol Standards.

A proper result for the standard test using a cylinder of this lot number would be the range of 0.075 to 0.085 g ethanol/210 L of vapor (g/100 ml of blood or g/210 L of end expiratory breath).

The Intoxilyzer® will print out the value of the standard test in 3 digits on Intoxilyzer® Test Records (Form 106-G or Form 106-I8000).

The number of cylinders sent to each location will be based on their need. The standard may be used until the date of expiration as indicated by the manufacturer's Certificate of Analysis.

Margaret A. Pearson, State Toxicologist

3 September 2009
Date Approved

DRYGAZ™ by CALGAZ

CERTIFICATE OF ANALYSIS
EBS - ETHANOL BREATH STANDARD

CMI
316 EAST NINTH STREET
OWENSBORO, KY 42303

INVOICE#: 34588894
PO#: 373328 RA 5145
CMI ITEM #: 340129
DATE: 03 Sep 2009

ANALYTICAL ACCURACY: +/- 2%
PRODUCT EXPIRATION: 03 Sep 2011

CALGAZ LOT#: 659358
ETHANOL IN NITROGEN

COMPONENT	PPM	(BrAC)
ETHANOL	208.4	(0.080)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
NIST TRACEABLE STANDARDS*	D751989	209.1

* CERTIFICATION TRACEABLE TO N.I.S.T. RGM ETHANOL STANDARDS

MANUFACTURED DATE: 03 Sep 2009

CALGAZ CYLINDER SIZE: 6D

APPROVED BY : Carolyn Christian

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reportsare on file and copies will be furnished upon request."

CALGAZ, Div. of Air Liquide Advanced Technologies U.S. LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410)228-6400 Fax: (410)228-4251